



Designers and Manufacturers of Mobile Health Clinics

ADI INC dba ADI Mobile Health

Code Standards Used During Manufacture of Mobile Health Clinic

- Chassis – by Ford (V-10 Gasoline) or Freightliner (Diesel) FVMSS or CVMSS
- Body – by Winnebago Industries Commercial Shell Division FVMSS or CVMSS
- RIVA 1192 – from ANSI/NFPA
 - For 12 VDC systems, batteries, battery charging
 - 120/240 VAC, 60 Htz, Single Phase AC System
 - Heating / Air Conditioning
 - LPG Heat Systems
 - Fire / Life / Safety
- ADI AC Electrical Systems – follows NEC which is from NFPA-70
- X-Ray / Radiation
 - Department of Health & Human Services, Food & Drug Administration Radiation Control Standards
 - Canadian Ministry of Health & Long Term Care – Canada
 - New York Public Health Law, Section 225, Part 1

ADI Vehicles and all features comply with all available Federal Motor Vehicle Safety Standards (FVMSS) and Canadian Vehicle Motor Safety Standards (CVMSS) as specified in US Code Title 49, Section 401