

ADI Mobile Clinic Maintenance Check List

DAILY TASKS:

1. Vehicle (walk-around counter clockwise):

- Log book: record date & mileage
- Check the Engine oil level
- Check the Coolant level
- Check the Windshield washer fluid level
- Check the Power steering fluid level
- Visual check of Engine
- Mirrors: clean & adjust
- Check the Wiper blades condition
- Clean side / rear view cameras
- Signal, Brake, Reverse, Running Lights working
- Check that the tire pressure is inflated to the recommended PSI
- Check the Tire tread condition
- Check the Exterior vehicle body condition
- Check the Drag Wheels
- Edges of Roof: check for tree damage
- AC cowlings: check for tree damage
- Awning: check for tree damage

2. Generator

- Log book: record date & hours
- Check the Generator Oil level - Hours to next oil change?
- Check the Coolant level
- Fan test switches (#1 & #2) to verify function
- Visual check of generator
- Error codes?

3. Mobile Clinic

- Power cord condition? Replace if frayed, cracked or cut. Replace Plug if cracked or broken.
- Check Step function
- Check Handrail condition
- Basement compartment latches function
- Check Potable & Waste water level s
- Check Battery voltage at monitor panel: #1,2,3 w/ charger turned "off"
- Check LPG tank level (if applicable)
- Sink faucet function: no drips?
- First Aid kit properly stocked?
- Fire Extinguishers in place and secure?
- Cabinets & drawers latching and functioning smoothly?
- Floors swept and clean?
- Windows latching properly?

4. Dental Equipment (if applicable)

- Vacuum chair traps cleaned?
- Vacuum system cleaner run through traps at assistant arm clean-out?
- Flush dental water lines 20 second between patients
- Vacuum pump trap cleaned?
- Ultrasonic cleaner drained?
- Empty dental water bottle and purge dry with air nightly if using a weekly water line treatment (commercial or 12 ml of bleach)
- Stools, carts, x-ray tie-downs and docks in good condition?

WEEKLY TASKS

- Visual check of battery fluid level and check with hydrometer, record in log book
- Check vacuum Hg separator level (if applicable)
- Dental / Medical water line system treatment (if applicable)
- Eyewash function checked?

MONTHLY TASKS

- Log book: record date & tire pressure
- Clean A/C filters: Log book: record date
- Flush out potable & waste tanks
- Treat potable water system & tank
- Check air compressor oil (if applicable)
- Check air compressor filter indicator
- Sterilizer filter checked
- Cycle UVL and note in logbook
- Pressure wash underside of vehicle to remove road salt
- Lubricate Auto-steps with Kwiikee Lube or equivalent.

BI-MONTHLY TASKS

- Dental water system bio-film monitoring program per manufacturer recommendation (if applicable)
- Autoclave test strip monitoring program (if applicable)
- Change generator engine oil every 300 hours, Log book: record date & Hours
- Lube leveling jacks with "Kwiikee" lube or equal

6 MONTHS

- Check engine air filter, Log book: record date
- Check Generator air filter

ANNUAL TASKS

- Change engine coolant per Ford or Freightliner recommendations, note in log book
- Change generator coolant with 50/50, note in log book
- Have basement compartment latches & key cylinders cleaned & lubricated.
- Inspect vehicle for underbody corrosion and road salt damage.
- Have Fire Extinguishers inspected and certified
- Have generator serviced at 2,500/5,000/7,500/10,000 Hours (Automotive equivalent: 150K/300K/450K/600K miles)

5 YEARS

- Have chamber autoclave checked by a qualified technician at 10,000 cycles or 5 years (if applicable).
- Have vehicle and all support systems surveyed.

ADI Mobile Health - FACTORY

19425 SW 89th Avenue
Tualatin, OR 97062 USA
www.adi-mobilehealth.com

Phone: +1-503-885-0886
Fax: +1-503-885-0417
tech@adi-mobilehealth.com
sales@adi-mobilehealth.com